Marketing Brochure TO PRINT_Layout 1 18/04/2017 16:40 Page 2

YOUR ESSENTIAL GUIDE TO SPIROMETRY

What is spirometry?

Spirometry is used to measure lung volumes and air flow. Alongside clinical assessment, it is an essential tool used in the diagnosis, assessment and monitoring of Chronic Obstructive Pulmonary Disease (COPD)¹, may contribute to the diagnosis of asthma and detect restrictive respiratory conditions.²



Diagnostic Spirometry: National Register of certified professionals and operators

As part of a phased introductory process starting in April 2017 all healthcare practitioners in England will be required to demonstrate that they are competent to perform and/or interpret spirometry and join a register of certified practitioners. The new scheme sets out how healthcare professionals performing and/or interpreting diagnostic spirometry should be trained, assessed and certified.


The key principles of the National register are:-

- Diagnostic spirometry must meet the Association for Respiratory Technology and Physiology (ARTP) standards
- Education and training must be flexible and accessible
- Assessment and verification process must include recognition of prior experience and competence

The register has three levels of competency assessment

- **Foundation**: those who have been assessed as competent to perform safe, accurate and reliable spirometry tests without interpretation
- Full: those who have been assessed as competent to perform and interpret spirometry in terms of physiological changes
- Interpretation Only: those who have been assessed as competent in interpretation only

How to become accredited on the National **Spirometry Register**



THE CERTIFICATION SCHEME TO RAISE STANDARDS OF SPIROMETRY FOR RESPIRATORY DIAGNOSIS:

Recognises the distinct and separate skills of performing spirometry, interpreting spirometry and diagnosing respiratory conditions where spirometry results form part of the whole picture.

... Training requirement identified following EPS assessment

- Enables practitioners who are experienced and confident in either performing and/or assessing spirometry and consider themselves competent to be assessed nced practitioner sc
- Supports practitioners who are less confident about their skills to access the appropriate training.
- Recognises that providing a good diagnostic spirometry service may take different forms in different places and is not prescriptive about the setting in which spirom-
- Will be phased in gradually up to March 2021.

FURTHER INFORMATION

https://www.pcc-cic.org.uk/article/quality-assured-diagnostic-spirometry http://www.artp.org.uk/en/spirometry/spiro-register/

What measurements are undertaken using spirometry?3

 Relaxed or slow vital capacity (VC) The volume of air that can be slowly expelled from the lung from maximal inspiration to maximum expiration

Forced vital capacity (FVC) The volume of air that can be forcibly expelled from

the lung from maximal inspiration to maximum expiration

Forced Expiratory Volume in 1 second (FEV₁)

The volume of air that can be forcibly expelled from maximum inspiration in the first second

FEV₁/FVC ratio

The FEV₁/FVC ratio is the FEV₁ expressed as a percentage of the FVC (or VC if that is greater). i.e. the proportion of the vital capacity exhaled in the first second. It distinguishes between a reduced FEV₁ due to restrictive lung volume and that due to obstruction. Obstruction is defined as an FEV₁/FVC ratio less than 70%


 Forced Expiratory Volume in 6 seconds (FEV₆)

The volume of air that can be forcibly expelled from maximum inspiration in six seconds.

This measurement is sometimes used as an alternative for FVC. Similarly FEV₁/FEV₆ is sometimes used instead of FEV₁/FVC.

Abnormal spirometry is divided into restrictive and obstructive ventilatory patterns.

- Restrictive patterns appear in conditions where the lung volume is reduced e.g. interstitial lung diseases, scoliosis. The FVC and FEV₁ are reduced proportionately
- Obstructive patterns appear when the airways are obstructed e.g. due to asthma or COPD. The FEV1 is reduced more than the FVC



5. Mark L Levy, Philip H Quanjer, Booker Rachel, Brendan G Cooper, Stephen Holmes & Iain R Small. Diagnostic Spirometry in Primary Care: Proposed standards for general practice compliant with American Thoracic Society and European Respiratory

Poorly performed spirometry is meaningless.

Who should undertake spirometry?

Spirometry should only be undertaken by healthcare professionals who are trained and competent (accredited) in performing (and ideally, interpreting) the tests.^{3,4,5} Regular updates and quality audits are fundamental to ensuring the quality of spirometry testing.

Accredited training courses include:-

Institution

Course

http://www.artp.org.uk/ Association for Respiratory Technology & Physiology

Association for Respiratory Technology & Physiology

The ARTP with the **British Thoracic Society** (BTS) offer a variety of training methods and an accreditation system to ensure acceptable standards of spirometry testing and interpretation.

https://www.educationforhealth.org/ **Education for Health**



Education for Health have a range of Spirometry courses written by experts, including workshops for those who simply need to feel more confident recording accurate measurements. The spirometry modules are developed with the Association of Respiratory Technology & Physiology (ARTP) and supported by the British Thoracic Society (BTS).

Types of spirometry testing⁴

- **Baseline testing** Used to investigate lung function where diagnosis has not been established.
- **Post-bronchodilator testing**
- o Investigative: To diagnose obstructive conditions where baseline spirometry shows an obstructive pattern
- o **Monitoring**: To monitor clinical progress in diagnosed asthma and COPD
- **Reversibility testing** May help to differentiate asthma from COPD.

What equipment is required to conduct spirometry?^{4,6}

- Spirometer (must meet ISO standard
- o Small hand-held meters which provide digital readings (but no visual display) are a cheap option which may be useful as a screening tool to identify people with abnormal readings who should be assessed by full diagnostic spirometry⁵
- One-way disposable mouthpieces and nose clips
- Bacterial and viral filters (selected patients with any risk of infection)
- Accurate height measures calibrated according to manufacturer's instructions
- Short-acting bronchodilators for reversibility testing and suitable means for delivery (volumatic/nebuliser)




Calibration, verification and maintenance of spirometry equipment³⁻⁶

Calibration of spirometry test equipment should be performed using a certificated 3 litre syringe and following the manufacturer's recommended procedures. For a device to be within calibration limits it must read +/- 3% of true. 4 Calibration should be verified prior to each clinic/session or after every 10th patient

maintenance.

Spirometers should be cleaned and service/ maintenance processes carried out regularly according to the manufacturer's instructions and in line with local and national guidance for infection control and equipment



Contraindications to spirometry testing³⁻⁶

Absolute

- Active infection e.g. AFB positive TB until treated for 2 weeks
- Conditions that may cause serious consequences to health if aggravated by forced expiration e.g. dissecting/unstable aortic aneurysm, pneumothorax, recent surgery (abdominal, thoracic, neurosurgery, eye surgery)

Relative

- Suspected respiratory infection in the last 4-6 weeks requiring antibiotics or steroids
- Undiagnosed chest symptoms e.g. haemoptysis
- Any condition which may be aggravated by forced expiration e.g. prior pneumothorax, history of myocardial infarction, stroke or embolism in the last 3 months, previous thoracic, abdominal or eye surgery
- Perforated ear drum
- Acute disorders such as nausea and vomiting
- Confusion, communication problems

COMMON ERRORS IN SPIROMETRY TESTING⁶

- Poor seal around mouthpiece Hesitation or false start
- Early termination of exhalation: a 'short blow' which has not achieved the full FVC
- Poor intake of breath

verified

- Poor forced expiratory effort Cough during procedure
- Incorrect data entered into the
- spirometer prior to testing Spirometer not calibrated and

the FEV₁ and FVC by the factors below⁶

Adjusting Caucasian reference values to other ethnic groups. To apply these, multiply

Population	FEV ₁	FVC
Hong Kong Chinese	1.0	1.0
Japanese American	0.89	-
Polynesian	0.9	0.9
North Indian and Pakistani	0.9	0.9
South Indian, African	0.87	0.87

Further Information for Patients http://patient.info/health/spirometry-leaflet

http://www.pcrs-uk.org

Society recommendations. A Primary Care Respiratory Society UK (PCRS-UK) document, in association with the Association for Respiratory Technology & Physiology (ARTP) and Education for Health. Prim Care Respir J. 2009; 18:130–147. 6. Spirometry in COPD Protocol. Primary Care Respiratory Society UK 2010. Available at https://www.pcrs-uk.org/resource/Nurse-tools/pcrs-uk-protocol-spirometry-copd-pdf

2. British Thoracic Society – Scottish Intercollegiate Guideline Network. British Guideline on the Management of Asthma. Thorax 2008;63 (Suppl 4): 1-121 Last Update October 2014.

4. A guide to performing quality assured diagnostic spirometry. 2013 Primary Care Commissioning. Available at http://www.pcc-cic.org.uk/article/guide-quality-assured-diagnostic-spirometry

1. National Institute for Health and Care Excellence. Management of chronic obstructive pulmonary disease (COPD) in adults in primary and secondary care (partial update) 2010 http://www.nice.org.uk/CG101

The guidance provided on this wall chart has been adapted from the following resources and publications:-

3. Spirometry PCRS-UK opinion Sheet Number 1, version 5. 2012. Available at https://www.pcrs-uk.org/resource/Opinion-sheets/spirometry-opinion-sheet

Available from https://www.brit-thoracic.org.uk/guidelines-and-quality-standards/asthma-guideline,



http://www.artp.org.uk/en/patient/lung-function-tests/pretest-info.cfm

PCRS-UK National Primary Care Conference 2017

Beyond the respiratory consultation: inspiring lifelong change

29-30th September 2017 **Telford International Centre**



Embedded in day-to-day clinical practice, this year's PCRS-UK conference will explore how we can work with patients in the consultation to help them bring about long term sustainable improvements not only in their respiratory condition but also their overall health and well being, through active participation in their care.

A holistic programme

The programme, designed by a multi-disciplinary team of primary, community and integrated care respiratory experts, will provide thought-provoking, interactive and inspiring keynote plenary presentations, patient-centred, essential clinical updates, key service development and commissioning sessions, real-world research presentations and practical workshops.

Conference partners





The Primary Care Respiratory Society UK are grateful to the following conference sponsors: AstraZeneca UK Limited, Boehringer Ingelheim Limited, Chiesi Limited, GlaxoSmithKline, Novartis Pharmaceuticals UK Limited, Johnson & Johnson and Pfizer Limited

For more information visit https://pcrs-uk.org/annual-conference

Registration

fees from

£150

PCRS-UK Membership - Join Today

Helping you to deliver high value patient centred respiratory care





- Membership emails and news alerts making it easy to keep up to date
- ✓ Huge savings on registration for our annual national primary care conference
- ✓ Professional development support including access to our clinical leadership programme
- Friendly community of like minded peers passionate about respiratory care



It pays to join if you are a respiratory lead for your practice or a respiratory health professional working in the community -

annual membership just £59 inclusive of VAT

The Primary Care Respiratory Society UK is grateful to its corporate supporters including AstraZeneca UK Ltd, Boehringer Ingelheim Ltd, Chiesi Ltd, Johnson & Johnson, Napp Pharmaceuticals, Novartis UK, Pfizer Ltd and TEVA UK Ltd for their financial support

Inspiring best practice in respiratory care

OPEN UP FOR YOUR ESSENTIAL CUIDE TO SPIROMETRY



programmes Clinical Leadership event for forthcoming https://pcrs-uk.org/clinical-leadership-programme



primary or community care setting. challenges of bringing about change in a clinicians who understand the realities and experienced clinical leaders - all practising Access to a multi disciplinary faculty of

network with like minded colleagues develop and practice a range of skills and Supportive and safe environment to

Who is member of PCRS-UK commissioning primary respiratory care and (or manager) involved in providing or Free to attend, for any health professional

longer term changes to care as well as equipping you to bring about workshops, relevant to your day to day work Interactive skills & knowledge based

respiratory community Things happen in your Equipping you to make

Respiratory clinical leadership programme

For more information see https://pcrs-uk.org/local-groups



- FREE PCRS-UK membership for leaders of affiliated local groups
- well as a network of support from other group leads Affiliation scheme offering enhanced credibility for your group as
 - development
- Annual meeting of group leaders supporting personal and group
- Support and resources to help you get started and develop your group

PCRS-UK is here to help you with:

- respiratory care in your area
- Develop and maintain links with other health professionals involved in
 - Support improved care in your area
 - Develop your personal and professional skills
 - Share best practice with, and learn from, your peers

If there is not one local to you, why not set up?

Visit www.pcrs-uk.org to find a PCRS-UK group near to you.

peers and benefit from a network of support. your area. Providing a forum to keep up to date, share best practice with A local respiratory group is the ideal way to bring colleagues together in



your area in respiratory care in Make a real difference

Local Respiratory Groups